



Methoden voor de bepaling van elementen



De volgende analysemethoden kunnen gebruikt worden voor het bepalen van elementen in water.

Voor de conservering en behandeling van watermonsters wordt verwezen naar WAC/I/A/010.

Voor drink- en grondwater dient geen ontsluiting te worden uitgevoerd.

Voor afval- en oppervlaktewater dient een ontsluiting te worden uitgevoerd conform WAC/III/B/001 of WAC/III/B/002.

	<ul style="list-style-type: none"> • ISO 11885:2007 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (WAC/III/B/010) • ISO 11969: 1996 Water quality – Determination of arsenic – Atomic absorption spectrometric method (hydride technique) (WAC/III/B/012) • ISO 17294-1: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines • ISO 17294-2: 2003 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (WAC/III/B/011) • NEN 6432:1993 Water – Bepaling van het gehalte aan arseen met behulp van atomaire absorptiespectrometrie (hydridegeneratietechniek). Ontsplitsing met salpeterzuur en zoutzuur. • EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace
arseen	<ul style="list-style-type: none"> • ISO 11885:2007 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (WAC/III/B/010) • EN 1233: 1996 Water quality – Determination of chromium – Atomic absorption spectrometric methods • ISO 17294-1: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines • ISO 17294-2: 2003 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (WAC/III/B/011) • EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace
chroom	<ul style="list-style-type: none"> • ISO 11885:2007 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (WAC/III/B/010) • ISO 8288: 1986 Water quality – Determination of cobalt, nickel, copper, zinc, cadmium and lead – Flame atomic absorption spectrometric methods • ISO 17294-1: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines • ISO 17294-2: 2003 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (WAC/III/B/011) • EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace
koper	<ul style="list-style-type: none"> • ISO 11885:2007 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (WAC/III/B/010) • ISO 17294-1: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines • ISO 17294-2: 2003 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (WAC/III/B/011) • EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace
lood	<ul style="list-style-type: none"> • ISO 11885:2007 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (WAC/III/B/010)

- ISO 8288: 1986 Water quality – Determination of cobalt, nickel, copper, zinc, cadmium and lead – Flame atomic absorption spectrometric methods
 - ISO 17294-1: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines
 - ISO 17294-2: 2003 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (WAC/III/B/011)
 - EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace
-
- ISO 11885:2007 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (WAC/III/B/010)
 - ISO 8288: 1986 Water quality – Determination of cobalt, nickel, copper, zinc, cadmium and lead – Flame atomic absorption spectrometric methods
 - ISO 17294-1: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines
 - ISO 17294-2: 2003 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (WAC/III/B/011)
 - EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace
-
- ISO 11885:2007 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (WAC/III/B/010)
 - ISO 17294-1: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines
 - ISO 17294-2: 2003 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (WAC/III/B/011)
 - EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace
-
- ISO 11885:2007 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (WAC/III/B/010)
 - ISO 8288: 1986 Water quality – Determination of cobalt, nickel, copper, zinc, cadmium and lead – Flame atomic absorption spectrometric methods
 - ISO 17294-1: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines
 - ISO 17294-2: 2003 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (WAC/III/B/011)
 - EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace
-
- ISO 11885:2007 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (WAC/III/B/010)
 - ISO 8288: 1986 Water quality – Determination of cobalt, nickel, copper, zinc, cadmium and lead – Flame atomic absorption spectrometric methods
 - ISO 5961: 1994 Water quality: Determination of cadmium by atomic absorption spectrometry
 - ISO 17294-1: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines
-

ijzer	<ul style="list-style-type: none"> • ISO 17294-2: 2003 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (WAC/III/B/011) • EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace
mangaan	<ul style="list-style-type: none"> • ISO 11885:2007 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (WAC/III/B/010) • ISO 17294-1: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines • ISO 17294-2: 2003 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (WAC/III/B/011) • EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace
seleen	<ul style="list-style-type: none"> • ISO 11885:2007 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (WAC/III/B/010) • ISO 17294-1: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines • ISO 17294-2: 2003 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (WAC/III/B/011) • NEN 6434:1993 Water – Bepaling van het gehalte aan seleen met behulp van atomaire absorptiespectrometrie (hydridegeneratietechniek). Ontsluiting met salpeterzuur en zoutzuur (WAC/III/B/012) • EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace
barium	<ul style="list-style-type: none"> • ISO 11885:2007 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (WAC/III/B/010) • ISO 17294-1: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines • ISO 17294-2: 2003 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (WAC/III/B/011)
antimoon	<ul style="list-style-type: none"> • ISO 11885:2007 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (WAC/III/B/010) • ISO 17294-1: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines • ISO 17294-2: 2003 Water quality – Application of inductively coupled

		plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (WAC/III/B/011)
	tin	<ul style="list-style-type: none"> • NEN 6433:1993 Water – Bepaling van het gehalte aan antimoon met behulp van atomaire absorptiespectrometrie (hydridegeneratietechniek). Ontsluiting met salpeterzuur en zoutzuur (WAC/III/B/012) • EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace
	aluminium	<ul style="list-style-type: none"> • ISO 11885:2007 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (WAC/III/B/010) • ISO 17294-1: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines • ISO 17294-2: 2003 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (WAC/III/B/011) • ISO 11885:2007 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (WAC/III/B/010) • ISO 17294-1: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines • ISO 17294-2: 2003 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (WAC/III/B/011) • ISO 12020: 1997 Water quality – Determination of aluminium – Atomic absorption spectrometric methods • EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace
	kobalt	<ul style="list-style-type: none"> • ISO 11885:2007 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (WAC/III/B/010) • ISO 8288: 1986 Water quality – Determination of cobalt, nickel, copper, zinc, cadmium and lead – Flame atomic absorption spectrometric methods • ISO 17294-1: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines • ISO 17294-2: 2003 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (WAC/III/B/011) • EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace
	molybdeen	<ul style="list-style-type: none"> • ISO 11885:2007 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (WAC/III/B/010) • ISO 17294-1: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines • ISO 17294-2: 2003 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (WAC/III/B/011) • EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace
	titaan	<ul style="list-style-type: none"> • ISO 11885:2007 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (WAC/III/B/010)

	<ul style="list-style-type: none"> • ISO 17294-1: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines • ISO 17294-2: 2003 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (WAC/III/B/011)
fosfor	<ul style="list-style-type: none"> • Zie WAC/III/D in functie van het matrixtype
natrium	<ul style="list-style-type: none"> • ISO 11885:2007 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (WAC/III/B/010) • ISO 17294-1: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines • ISO 17294-2: 2003 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (WAC/III/B/011) • ISO 14911:1998 Water quality - Determination of dissolved Li⁺, Na⁺, NH₄⁺, K⁺, Mn²⁺, Ca²⁺, Mg²⁺, Sr²⁺ and Ba²⁺ using ion chromatography - Method for water and waste water
magnesium	<ul style="list-style-type: none"> • ISO 11885:2007 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (WAC/III/B/010) • ISO 17294-1: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines • ISO 17294-2: 2003 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (WAC/III/B/011) • ISO 14911:1998 Water quality - Determination of dissolved Li⁺, Na⁺, NH₄⁺, K⁺, Mn²⁺, Ca²⁺, Mg²⁺, Sr²⁺ and Ba²⁺ using ion chromatography - Method for water and waste water
kalium	<ul style="list-style-type: none"> • ISO 11885:2007 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (WAC/III/B/010) • ISO 17294-1: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines • ISO 17294-2: 2003 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (WAC/III/B/011) • ISO 14911:1998 Water quality - Determination of dissolved Li⁺, Na⁺, NH₄⁺, K⁺, Mn²⁺, Ca²⁺, Mg²⁺, Sr²⁺ and Ba²⁺ using ion chromatography - Method for water and waste water
calcium	<ul style="list-style-type: none"> • ISO 11885:2007 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (WAC/III/B/010) • ISO 17294-1: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines • ISO 17294-2: 2003 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (WAC/III/B/011) • ISO 14911:1998 Water quality - Determination of dissolved Li⁺, Na⁺, NH₄⁺, K⁺, Mn²⁺, Ca²⁺, Mg²⁺, Sr²⁺ and Ba²⁺ using ion chromatography - Method for water and waste water

Voor de bepaling van **kwik** is het toepassen van een ontsluiting afhankelijk van de conserverings- en bepalingstechniek **en het type water**.

Voor de conservering en behandeling van watermonsters wordt verwezen naar WAC/I/A/010.

Bij de bepaling van kwik met CV-AAS en CV-AFS (WAC/III/B/014) en BrCl als conservering reagens dient voor drink- en grondwater geen ontsluiting te worden uitgevoerd. Voor afval- en oppervlaktewater dient een ontsluiting te worden uitgevoerd, tenzij de nodige gegevens beschikbaar zijn die aantonen dat het type afvalwater dat door het laboratorium wordt geanalyseerd gelijkwaardige resultaten geeft zonder ontsluiting bij verhoogde temperatuur.

Volgende ontsluitingsmethoden kunnen worden toegepast:

- WAC/III/B/001 Ontsleuteling voor de bepaling van geselecteerde elementen in water – salpeterzuurontsluiting
- WAC/III/B/002 Ontsleuteling voor de bepaling van geselecteerde elementen in water – aqua regia ontsluiting
- Ontsleuteling met BrCl oplossing bij kamertemperatuur gedurende min. 24u volgens ISO/FDIS 12846 § 7.4 (kT)
- Ontsleuteling met BrCl reagens bij verhoogde temperatuur, BrCl (hT)¹

Bij de bepaling van kwik met CV-AAS en CV-AFS (WAC/III/B/014) en kaliumdichromaat als conservering reagens dient zowel drink-, grond-, afval- en oppervlaktewater ontsloten te worden.

Volgende ontsluitingsmethoden kunnen worden toegepast:

- WAC/III/B/002 Ontsleuteling voor de bepaling van geselecteerde elementen in water – aqua regia ontsluiting
- Ontsleutingsmethode met kalium permanganaat/ kalium perroxodisulfaat
 - Breng 100 ml monster geconserveerd met HNO₃ en K₂Cr₂O₇ (0.05%), in een ontsluitingsrecipiënt
 - Voeg 15 ml kalium permanganaat oplossing (50 g KMnO₄/liter), 1 ml HNO₃ en 1 ml H₂SO₄ toe.
 - Laat de oplossing 15 min. staan, en voeg 10 ml kalium perroxodisulfaat oplossing (40 g K₂S₂O₈/liter) toe
 - Plaats het recipiënt in een verwarmingsblok of waterbad bij 95°C gedurende 2uur
 - Indien nodig, voeg bijkomend kalium permanganaat oplossing toe
 - Laten afkoelen, en aanlengen tot gewenst volume.
- Ontsleuteling met BrCl oplossing bij kamertemperatuur gedurende min. 24u volgens ISO/FDIS 12846 § 7.4 (kT)
- Ontsleuteling met BrCl reagens bij verhoogde temperatuur, BrCl (hT)¹

Bij de bepaling van Hg met ICP-MS dient onafhankelijk van de conservering voor drink- en grondwater geen ontsluiting te worden uitgevoerd. **Bijkomend** dient bij de analyse AuCl₃ te worden toegevoegd aan zowel standaarden als monsters om geheugeneffecten in de

¹ Method 1631, Revision E: Mercury in Water by Oxidation, Purge and Trap, and Cold Vapor Atomic Fluorescence Spectrometry, August 2002.

verstuiverkamer te minimaliseren². Voor afval- en oppervlaktewater dient een ontsluiting te worden uitgevoerd.

Volgende ontsluitingsmethoden kunnen worden toegepast:

- WAC/III/B/001 Ontsleuteling voor de bepaling van geselecteerde elementen in water – salpeterzuurontsluiting
- WAC/III/B/002 Ontsleuteling voor de bepaling van geselecteerde elementen in water – aqua regia ontsleuteling

De volgende analysemethoden kunnen gebruikt worden voor de bepaling van kwik in water.

-
- | | |
|------|--|
| Kwik | <ul style="list-style-type: none">• ISO/FDIS 12846:2011 Water quality : Determination of mercury - Method using atomic absorption spectrometry (AAS) with and without enrichment• ISO 17852:2006 Water quality : Determination of mercury - Method using atomic fluorescence spectrometry• EN 12338: 1998 Water quality: Determination of mercury – enrichment methods by amalgamation• EPA 200.8 Determination of trace elements in waters and wastes by inductively coupled plasma- mass spectrometry.• WAC/III/B/011 Bepaling van elementen met inductief gekoppeld plasma massa spectrometrie (ICP-MS)• ISO 5666: 1999 Water quality: Determination of mercury• EN 1483: 2007 Water quality – Determination of mercury |
|------|--|
-

² B.T. Sturman, *Comment on 'Determination of mercury in potable water by ICP-MS using gold as stabilising agent*, J. Anal. At. Spectrom., 2000, **15**, 1512.