

## Methoden voor de bepaling van elementen

Deze procedure vervangt de procedure CMA/2/I/B van juni 2014.

Voor de conservering en bewaring van watermonsters worden verwezen naar CMA/1/B.

## MATRICES: GRONDWATER, ELUATEN, DESTRUCTIEVLOEISTOFFEN

### ELEMENTEN

Voor grondwater en eluat dient geen ontsluiting te worden uitgevoerd. De analyses worden uitgevoerd op de gefiltreerde ( $0.45 \mu\text{m}$ ) monsters. Enkel indien een neerslag wordt waargenomen, dient een ontsluiting te worden uitgevoerd conform CMA/2/I/A.6.1 of CMA/2/I/A.6.3.

Voor bodem, vaste en pasteuze afvalstoffen dient een destructie te worden uitgevoerd conform CMA/2/II/A.3.

Voor materialen die als meststof/bodemverbeterend middel worden aangewend, dient de destructie te worden uitgevoerd conform CMA/2/IV/6 met uitzondering van vloeibare monsters met een droge stofgehalte < 2%. Deze worden behandeld als een afvalwater en ontsloten conform WAC/III/B/002.

Voor olie dient een destructie te worden uitgevoerd conform CMA/2/III/F.

De volgende analysetechnieken kunnen gebruikt worden voor de bepaling van de elementen in grondwater, eluat, destructievloeistoffen:

- |          |  |
|----------|--|
| antimoon | <ul style="list-style-type: none"><li>• NBN EN ISO 11885:2009 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (ISO 11885:2007) (CMA/2/I/B.1)</li><li>• NBN EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace (ISO 15586:2003) (CMA/2/I/B.2)</li><li>• NBN EN ISO 17294-1:2006 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines (ISO 17294-1:2004)</li><li>• NBN EN ISO 17294-2:2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (ISO 17294-2:2003) (CMA/2/I/B.5)</li><li>• CMA/2/I/B.6 Bepaling van antimoon, arseen en seleen met hydride-atomaire absorptie spectrometrie</li><li>• ISO 17378-1:2014 Water quality - Determination of arsenic and antimony - Part 1: Method using hydride generation atomic fluorescence spectrometry (HG-AFS)</li><li>• ISO 17378-2:2014 Water quality - Determination of arsenic and antimony - Part 2: Method using hydride generation atomic absorption spectrometry (HG-AAS)</li><li>• ISO 20280:2007 Soil quality -- Determination of arsenic, antimony and selenium in aqua regia soil extracts with electrothermal or hydride-generation atomic absorption spectrometry</li></ul> |
| arseen   | <ul style="list-style-type: none"><li>• NBN EN ISO 11885:2009 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (ISO 11885:2007) (CMA/2/I/B.1)</li><li>• NBN EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace (ISO 15586:2003) (CMA/2/I/B.2)</li><li>• NBN EN ISO 17294-1:2006 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines (ISO 17294-1:2004)</li><li>• NBN EN ISO 17294-2:2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (ISO 17294-2:2003) (CMA/2/I/B.5)</li><li>• CMA/2/I/B.6 Bepaling van antimoon, arseen en seleen met hydride-atomaire absorptie spectrometrie</li><li>• ISO 17378-1:2014 Water quality - Determination of arsenic and antimony - Part 1: Method using hydride generation atomic fluorescence spectrometry (HG-AFS)</li><li>• ISO 17378-2:2014 Water quality - Determination of arsenic and antimony - Part 2: Method using hydride generation atomic absorption spectrometry (HG-AAS)</li><li>• ISO 20280:2007 Soil quality -- Determination of arsenic, antimony and selenium in aqua regia soil extracts with electrothermal or hydride-generation atomic absorption spectrometry</li></ul> |

	<p>elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (ISO 11885:2007) (CMA/2/I/B.1)</p> <ul style="list-style-type: none"> <li>• NBN EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace (ISO 15586:2003) (CMA/2/I/B.2)</li> <li>• NBN EN ISO 17294-1:2006 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines (ISO 17294-1:2004)</li> <li>• NBN EN ISO 17294-2:2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (ISO 17294-2:2003) (CMA/2/I/B.5)</li> <li>• CMA/2/I/B.6 Bepaling van antimoon, arseen en seleen met hydride-atomaire absorptie spectrometrie</li> <li>• ISO 17378-1:2014 Water quality - Determination of arsenic and antimony - Part 1: Method using hydride generation atomic fluorescence spectrometry (HG-AFS)</li> <li>• ISO 17378-2:2014 Water quality - Determination of arsenic and antimony - Part 2: Method using hydride generation atomic absorption spectrometry (HG-AAS)</li> <li>• NEN 6432:1993 Water – Bepaling van het gehalte aan arseen met behulp van atomaire absorptiespectrometrie (hydridegeneratie-techniek). Ontsluiting met salpeterzuur en zoutzuur.</li> <li>• ISO 20280:2007 Soil quality -- Determination of arsenic, antimony and selenium in aqua regia soil extracts with electrothermal or hydride-generation atomic absorption spectrometry</li> </ul>
barium	<ul style="list-style-type: none"> <li>• NBN EN ISO 11885:2009 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (ISO 11885:2007) (CMA/2/I/B.1)</li> <li>• NBN EN ISO 17294-1:2006 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines (ISO 17294-1:2004)</li> <li>• NBN EN ISO 17294-2:2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (ISO 17294-2:2003) (CMA/2/I/B.5)</li> </ul>
cadmium	<ul style="list-style-type: none"> <li>• NBN EN ISO 11885:2009 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (ISO 11885:2007) (CMA/2/I/B.1)</li> <li>• NBN EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace (ISO 15586:2003) (CMA/2/I/B.2)</li> <li>• NBN EN ISO 17294-1:2006 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines (ISO 17294-1:2004)</li> <li>• NBN EN ISO 17294-2:2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (ISO 17294-2:2003) (CMA/2/I/B.5)</li> <li>• ISO 8288: 1986 Water quality – Determination of cobalt, nickel, copper, zinc, cadmium and lead – Flame atomic absorption spectrometric methods</li> </ul>

	<ul style="list-style-type: none"> <li>• ISO 5961: 1994 Water quality: Determination of cadmium by atomic absorption spectrometry</li> </ul>
calcium	<ul style="list-style-type: none"> <li>• NBN EN ISO 11885:2009 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (ISO 11885:2007) (CMA/2/I/B.1)</li> <li>• NBN EN ISO 17294-1:2006 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines (ISO 17294-1:2004)</li> <li>• NBN EN ISO 17294-2:2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (ISO 17294-2:2003) (CMA/2/I/B.5)</li> </ul>
chroom	<ul style="list-style-type: none"> <li>• NBN EN ISO 11885:2009 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (ISO 11885:2007) (CMA/2/I/B.1)</li> <li>• NBN EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace (ISO 15586:2003) (CMA/2/I/B.2)</li> <li>• NBN EN ISO 17294-1:2006 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines (ISO 17294-1:2004)</li> <li>• NBN EN ISO 17294-2:2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (ISO 17294-2:2003) (CMA/2/I/B.5)</li> <li>• NBN EN 1233: 1996 Water quality – Determination of chromium – Atomic absorption spectrometric methods</li> <li>• CEN/TS 16188:2012 Sludge, treated biowaste and soil – Determinations of elements in aqua regia and nitric acid digests – Flame atomic absorption spectrometry method (FAAS)</li> </ul>
fosfor	<ul style="list-style-type: none"> <li>• NBN EN ISO 11885:2009 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (ISO 11885:2007) (CMA/2/I/B.1)</li> <li>• NBN EN ISO 6878: 2004 Water quality – Determination of phosphorus – Ammonium molybdate spectrometric method (ISO 6778:2004)</li> <li>• NBN EN ISO 17294-1:2006 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines (ISO 17294-1:2004)</li> <li>• NBN EN ISO 17294-2:2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (ISO 17294-2:2003) (CMA/2/I/B.5)</li> <li>• NBN EN ISO 15681-1:2005 Water quality – Determination of orthophosphate and total phosphorus contents by flow analysis (FIA and CFA) – Part 1: Method by flow injection analysis (FIA) (ISO 15681-1:2003)</li> <li>• NBN EN ISO 15681-2:2005 Water quality – Determination of orthophosphate and total phosphorus contents by flow analysis (FIA and CFA) – Part 2: Method by continuous flow analysis (CFA) (ISO 15681-2:2003)</li> </ul>
kalium	<ul style="list-style-type: none"> <li>• NBN EN ISO 11885:2009 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry</li> </ul>

	(ICP-OES) (ISO 11885:2007) (CMA/2/I/B.1)
•	NBN EN ISO 17294-1:2006 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines (ISO 17294-1:2004)
•	NBN EN ISO 17294-2:2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (ISO 17294-2:2003) (CMA/2/I/B.5)
•	NBN EN ISO 11885:2009 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (ISO 11885:2007) (CMA/2/I/B.1)
•	NBN EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace (ISO 15586:2003) (CMA/2/I/B.2)
•	NBN EN ISO 17294-1:2006: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines (ISO 17294-1:2004)
•	NBN EN ISO 17294-2:2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (ISO 17294-2:2003) (CMA/2/I/B.5)
•	ISO 8288: 1986 Water quality – Determination of cobalt, nickel, copper, zinc, cadmium and lead – Flame atomic absorption spectrometric methods
•	CEN/TS 16188:2012 Sludge, treated biowaste and soil – Determinations of elements in aqua regia and nitric acid digests – Flame atomic absorption spectrometry method (FAAS)
•	NBN EN ISO 11885:2009 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (ISO 11885:2007) (CMA/2/I/B.1)
•	NBN EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace (ISO 15586:2003) (CMA/2/I/B.2)
•	NBN EN ISO 17294-1:2006 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines (ISO 17294-1:2004)
•	NBN EN ISO 17294-2:2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (ISO 17294-2:2003) (CMA/2/I/B.5)
•	ISO 8288: 1986 Water quality – Determination of cobalt, nickel, copper, zinc, cadmium and lead – Flame atomic absorption spectrometric methods
•	CEN/TS 16188:2012 Sludge, treated biowaste and soil – Determinations of elements in aqua regia and nitric acid digests – Flame atomic absorption spectrometry method (FAAS)
•	NBN EN ISO 11885:2009 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (ISO 11885:2007) (CMA/2/I/B.1)
•	NBN EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace (ISO 15586:2003) (CMA/2/I/B.2)

	<ul style="list-style-type: none"> <li>• NBN EN ISO 17294-1:2006 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines (ISO 17294-1:2004)</li> <li>• NBN EN ISO 17294-2:2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (ISO 17294-2:2003) (CMA/2/I/B.5)</li> <li>• ISO 8288: 1986 Water quality – Determination of cobalt, nickel, copper, zinc, cadmium and lead – Flame atomic absorption spectrometric methods</li> </ul>
magnesium	<ul style="list-style-type: none"> <li>• NBN EN ISO 11885:2009 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (ISO 11885:2007) (CMA/2/I/B.1)</li> <li>• NBN EN ISO 17294-1:2006 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines (ISO 17294-1:2004)</li> <li>• NBN EN ISO 17294-2:2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (ISO 17294-2:2003) (CMA/2/I/B.5)</li> </ul>
mangaan	<ul style="list-style-type: none"> <li>• NBN EN ISO 11885:2009 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (ISO 11885:2007) (CMA/2/I/B.1)</li> <li>• NBN EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace (ISO 15586:2003) (CMA/2/I/B.2)</li> <li>• NBN EN ISO 17294-1:2006 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines (ISO 17294-1:2004)</li> <li>• NBN EN ISO 17294-2:2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (ISO 17294-2:2003) (CMA/2/I/B.5)</li> <li>• CEN/TS 16188:2012 Sludge, treated biowaste and soil – Determinations of elements in aqua regia and nitric acid digests – Flame atomic absorption spectrometry method (FAAS)</li> </ul>
molybdeen	<ul style="list-style-type: none"> <li>• NBN EN ISO 11885:2009 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (ISO 11885:2007) (CMA/2/I/B.1)</li> <li>• NBN EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace (ISO 15586:2003) (CMA/2/I/B.2)</li> <li>• NBN EN ISO 17294-1:2006 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines (ISO 17294-1:2004)</li> <li>• NBN EN ISO 17294-2:2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (ISO 17294-2:2003) (CMA/2/I/B.5)</li> </ul>
nikkel	<ul style="list-style-type: none"> <li>• NBN EN ISO 11885:2009 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (ISO 11885:2007) (CMA/2/I/B.1)</li> <li>• NBN EN ISO 15586:2003 Water quality – Determination of trace</li> </ul>

	elements using atomic absorption spectrometry with graphite furnace (ISO 15586:2003) (CMA/2/I/B.2)
	• NBN EN ISO 17294-1:2006 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines (ISO 17294-1:2004)
	• NBN EN ISO 17294-2:2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (ISO 17294-2:2003) (CMA/2/I/B.5)
	• ISO 8288: 1986 Water quality – Determination of cobalt, nickel, copper, zinc, cadmium and lead – Flame atomic absorption spectrometric methods
	• CEN/TS 16188:2012 Sludge, treated biowaste and soil – Determinations of elements in aqua regia and nitric acid digests – Flame atomic absorption spectrometry method (FAAS)
seleen	• NBN EN ISO 11885:2009 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (ISO 11885:2007) (CMA/2/I/B.1)
seleen	• NBN EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace (ISO 15586:2003) (CMA/2/I/B.2)
seleen	• NBN EN ISO 17294-1:2006 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines (ISO 17294-1:2004)
seleen	• NBN EN ISO 17294-2:2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (ISO 17294-2:2003) (CMA/2/I/B.5)
seleen	• CMA/2/I/B.6 Bepaling van antimoon, arseen en seleen met hydride-atomaire absorptie spectrometrie
seleen	• ISO/TS 17379-1:2013 Water quality - Determination of selenium - Part 1: Method using hydride generation atomic fluorescence spectrometry (HG-AFS)
seleen	• ISO/TS 17379-2:2013 Water quality - Determination of selenium - Part 2: Method using hydride generation atomic absorption spectrometry (HG-AAS)
seleen	• ISO 20280:2007 Soil quality -- Determination of arsenic, antimony and selenium in aqua regia soil extracts with electrothermal or hydride-generation atomic absorption spectrometry
thallium	• NBN EN ISO 11885:2009 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (ISO 11885:2007) (CMA/2/I/B.1)
thallium	• NBN EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace (ISO 15586:2003) (CMA/2/I/B.2)
thallium	• NBN EN ISO 17294-1:2006 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines (ISO 17294-1:2004)
thallium	• NBN EN ISO 17294-2:2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (ISO 17294-2:2003) (CMA/2/I/B.5)

tin	<ul style="list-style-type: none"><li>• NBN EN ISO 11885:2009 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (ISO 11885:2007) (CMA/2/I/B.1)</li><li>• NBN EN ISO 17294-1:2006 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines (ISO 17294-1:2004)</li><li>• NBN EN ISO 17294-2:2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (ISO 17294-2:2003) (CMA/2/I/B.5)</li></ul>
vanadium	<ul style="list-style-type: none"><li>• NBN EN ISO 11885:2009 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (ISO 11885:2007) (CMA/2/I/B.1)</li><li>• NBN EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace (ISO 15586:2003) (CMA/2/I/B.2)</li><li>• NBN EN ISO 17294-1:2006 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines (ISO 17294-1:2004)</li><li>• NBN EN ISO 17294-2:2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (ISO 17294-2:2003) (CMA/2/I/B.5)</li></ul>
zink	<ul style="list-style-type: none"><li>• NBN EN ISO 11885:2009 Water quality — Determination of selected elements by inductively coupled plasma optical emission spectrometry (ICP-OES) (ISO 11885:2007) (CMA/2/I/B.1)</li><li>• NBN EN ISO 15586:2003 Water quality – Determination of trace elements using atomic absorption spectrometry with graphite furnace (ISO 15586:2003) (CMA/2/I/B.2)</li><li>• NBN EN ISO 17294-1:2006 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines (ISO 17294-1:2004)</li><li>• NBN EN ISO 17294-2:2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (ISO 17294-2:2003) (CMA/2/I/B.5)</li><li>• ISO 8288: 1986 Water quality – Determination of cobalt, nickel, copper, zinc, cadmium and lead – Flame atomic absorption spectrometric methods</li><li>• CEN/TS 16188:2012 Sludge, treated biowaste and soil – Determinations of elements in aqua regia and nitric acid digests – Flame atomic absorption spectrometry method (FAAS)</li></ul>

**KWIK**

Voor de bepaling van kwik in grondwater en eluatien is het toepassen van een ontsluiting afhankelijk van de conserverings- en bepalingstechniek.

Bij de bepaling van kwik met CV-AAS en CV-AFS (CMA/2/I/B.3) en BrCl als conservering reagens dient voor grondwater en eluatien geen ontsluiting te worden uitgevoerd.

Bij de bepaling van kwik met CV-AAS en CV-AFS (CMA/2/I/B.3) en kaliumdichromaat als conservering reagens dient zowel grondwater als de eluatien ontsloten te worden.

Volgende ontsluitingsmethoden kunnen worden toegepast:

- CMA/2/I/A.6.1 Ontsleuteling voor de bepaling van geselecteerde elementen in water – aqua regia ontsleuteling
- Ontsleutelingsmethode met kalium permanganaat/ kalium perroxidisulfaat
  - Breng 100 ml monster geconserveerd met  $\text{HNO}_3$  en  $\text{K}_2\text{Cr}_2\text{O}_7$  (0.05%), in een ontsleutelingsrecipiënt
  - Voeg 15 ml kalium permanganaat oplossing (50 g  $\text{KMnO}_4$ /liter), 1 ml  $\text{HNO}_3$  en 1 ml  $\text{H}_2\text{SO}_4$  toe.
  - Laat de oplossing 15 min. staan, en voeg 10 ml kalium perroxidisulfaat oplossing (40 g  $\text{K}_2\text{S}_2\text{O}_6$ /liter) toe
  - Plaats het recipiënt in een verwarmingstoestel (bv. blok of waterbad) bij 95°C gedurende 2 uur
  - Indien nodig, voeg bijkomend kalium permanganaat oplossing toe
  - Laten afkoelen, en aanlengen tot gewenst volume.
- Ontsleuteling met  $\text{BrCl}$  oplossing bij kamertemperatuur gedurende min. 24 u volgens ISO 12846 § 7.4 (kT)
- Ontsleuteling met  $\text{BrCl}$ -reagens bij verhoogde temperatuur,  $\text{BrCl}$  (hT)<sup>1</sup>

Bij de bepaling van Hg met ICP-MS dient onafhankelijk van de conservering voor grondwater en eluatien geen ontsleuteling te worden uitgevoerd.

Voor bodem, vaste en pasteuze afvalstoffen dient een destructie te worden uitgevoerd conform CMA/2/II/A.3.

Voor materialen die als bodemverbeterende middel/meststof worden aangewend, dient de destructie te worden uitgevoerd conform CMA/2/IV/6.

De volgende analysetechnieken kunnen gebruikt worden voor de bepaling van kwik in grondwater, eluatien en destructievloeistoffen:

Kwik	<ul style="list-style-type: none"> <li>• NBN EN ISO 12846:2012 Water quality : Determination of mercury - Method using atomic absorption spectrometry (AAS) with and without enrichment (ISO 12846:2012) (CMA/2/I/B.3)</li> <li>• NBN EN ISO 17852:2008 Water quality -- Determination of mercury -- Method using atomic fluorescence spectrometry (ISO 17852:2006) (CMA/2/I/B.3)</li> <li>• <b><u>EN 12338: 1998 Water quality: Determination of mercury — enrichment</u></b></li> </ul>

<sup>1</sup> EPA Method 1631, Revision E: Mercury in Water by Oxidation, Purge and Trap, and Cold Vapor Atomic Fluorescence Spectrometry, August 2002.

**methods by amalgamation**

- NBN EN ISO 17294-1: 2006 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 1: General guidelines (ISO 17294-1:2004)
- NBN EN ISO 17294-2: 2004 Water quality – Application of inductively coupled plasma mass spectrometry (ICP-MS) – Part 2: Determination of 62 elements (ISO 17294-2:2003) (CMA/2/I/B.5)
- EPA 200.8:2007 Determination of trace elements in waters and wastes by inductively coupled plasma- mass spectrometry.
- ~~ISO 5666: 1999 Water quality: Determination of mercury (CMA/2/I/B.3)~~
- ~~EN 1483: 2007 Water quality - Determination of mercury (CMA/2/I/B.3)~~
- EPA 7473:1998 Mercury in solids and solutions by thermal decomposition, amalgamation, and atomic absorption spectrophotometry